


CCM.FF-K4, EURAMET.M.FF-K4.b and SIM.M.FF-K4

MEASURAND:

Volume

NOMINAL VALUE: 20 L

Degrees of equivalence D_i and expanded uncertainty U_i at approximately 95% level of confidence, expressed in 10^{-6} 

Red diamonds : participants in CCM.FF-K4

Blue circles : participants in EURAMET.M.FF-K4.b

Green triangles : participants in SIM.M.FF-K4

 $U_{BSJ} = 631 \times 10^{-6}$ $D_{LACOMET} = -235 \times 10^{-6}$


CCM.FF-K4 and SIM.M.FF-K4

MEASURAND: Volume

NOMINAL VALUE: 100 mL

TRANSFER STANDARD: TS 03.04.04

Degrees of equivalence D_i and expanded uncertainty U_i at approximately a 95% level of confidence



Red diamonds : participants in CCM.FF-K4

Green triangles : participants in SIM.M.FF-K4

$D_{BSJ} = 10601 \times 10^{-6}$

$U_{BSJ} = 721 \times 10^{-6}$