

INTERLABORATORY MASS COMPARISON BETWEEN LABORATORIES BELONGING TO SIM –SUB-REGIONS COORDINATED BY CENAM (SIM.7.31a & SIM.7.31b)

Jorge Nava¹, Luis O. Becerra¹, Tweedsmuir Mitchell², Mairy Sanz², Olman Ramos³, Sandra Rodríguez³, Gerson Vallejos⁴, Marco Bautista⁴, Francisco García⁵, Fernando Leyton⁵, Aldo Quiroga⁶, Juan G. Rodríguez⁶, Ruben Ramirez⁷, Arnoldo Florencio⁷

¹CENAM–México, ²BSJ–Jamaica, ³LACOMET–Costa Rica,

⁴IBMETRO–Bolivia, ⁵CESMEC–Chile, ⁶INDECOPI–Peru, ⁷INTN–Paraguay⁷

Tel: (+52) 442 2 11 05 00, Fax: (+52) 442 2 11 05 68, e-mail: jnava@cenam.mx, lbecerra@cenam.mx

Abstract: A round robin comparison in mass measurements between SIM member countries was carried out during the period April to November 2005. CENAM acted as pilot laboratory.

Six travelling standards with the following nominal values: 2 kg, 1 kg, 200 g, 50 g, 1 g and 200 mg were circulated. These travelling standards complied with the accuracy recommended for class E₂ [1,2]. The results obtained are represented in this report.

1. INTRODUCTION

A meeting of the technical contacts of SIM MWG 7 was held in Rio de Janeiro, Brazil in December 2004. At this meeting, subsequent to a proposal by the BSJ, planning was commenced for a mass comparison between SIM member countries in which at least one country from each sub region should participate. CENAM accepted the role as pilot laboratory for the mass comparison, as it had taken part in the key comparisons of the CCM of the CIPM.

The results of a comparison of six travelling standards among laboratories in the SIM region are presented in this report.

This program was coordinated by CENAM (Centro Nacional de Metrología), México. The travelling standards used are: 2 kg, 1 kg, 200 g, 50 g, 1 g and 200 mg, all of them are made of non-magnetic stainless steel.

The measurements in this comparison were carried out from April 2005 to December 2005. The CENAM contributed the travelling standards and supplied their reference values.

The density, the magnetic susceptibility, permanent magnetization, and conventional mass of all travelling standards except the density of the 200 mg were determined. A visual comparison of the surface roughness against roughness

standards proved that the travelling standards complied with the accuracy class of OIML E₂ [1,2].

The comparison protocol as well as the volume data were included in the travel container in which the standards were transported.

The SIM identification for this comparison is: **SIM.7.31a** (1 kg) and **SIM.7.31b** (2 kg, 200 g, 50 g, 1 g y 200 mg).

2. AIM OF THE PROGRAM

The aim of this comparison is to give confidence of the technical capacity of the SIM members and work in the mutual recognition agreements within the SIM and at the international level. On the other hand this comparison gives objective evidence about the technical competence of the laboratories, and it assists in identifying opportunities to improve the metrological assurance systems.

One of the problems in organizing comparisons, in which different countries are involved, is that each of them has different necessities and different capabilities; this can be seen in the wide range of uncertainty reported by the participants.

3. PARTICIPANTS

Table 1 shows the seven participating laboratories of the SIM sub-regions.

Table 1: Participating Laboratories

Laboratory	Acronym	Country	SIM Sub region
Centro Nacional de Metrología	CENAM/ Pilot laboratory	México	NORAMET
Bureau of Standards, Jamaica	BSJ	Jamaica	CARIMET
Laboratorio Costarricense de Metrología	LACOMET	Costa Rica	CAMET
Instituto Boliviano de Metrología	IBMETRO	Bolivia	ANDIMET
Centro de Estudios, Medición y Certificación de Calidad	CESMEC	Chile	SURAMET
Instituto Nacional de Defensa de la Competencia y de la Protección de la Propiedad Intelectual.	INDECOPI	Peru	ANDIMET
Instituto Nacional de Tecnología, Normalización y Metrología	INTN	Paraguay	SURAMET

4. DESIGN OF THE PROGRAM AND TIME SCHEDULE

The program was designed according to the guidelines for CIPM (Comité International des Poids et Measures) key comparisons [3] and were used six travelling standards of 2 kg, 1 kg, 200 g, 50 g, 1 g and 200 mg) were used. These standards comply with the requirements of class E₂ of the International Recommendation OIML R111 [1,2]. The travelling standards were circulated in only one petal of SIM sub-regions. As pilot laboratory, CENAM determined the conventional mass of the travelling standards at the beginning and the end of the comparison.


The transportation of the travelling standards to the next participant was done by hand in order to avoid any contamination or damage.

Table 2 shows the measurements scheduled and the starting dates of each laboratory.

Table 2:- Sequence of the measurements

INSTITUTE	PERIOD OF MEASUREMENTS
CENAM	April-May 2005
BSJ	May 2005
LACOMET	June 2005
IBMETRO	August 2005
INDECOPI	September 2005
INTN	October 2005
CESMEC	November 2005
CENAM	December 2005-January 2006

Figure 1 shows the transportation sequence and measurements of the travelling standards.



5. REPORTING BY PARTICIPANT

The measurement results were sent to the pilot laboratory in a final report where a list of the equipment used as balances and environmental conditions were included, besides the reference standard used in order to see the traceability on each laboratory.

6. STABILITY OF THE TRAVELLING STANDARDS

The pilot laboratory (CENAM) monitored the stability of the travelling standards during a period of 2 months before beginning of the laboratory measurements. No significant instability has been found, so that, the conventional mass values of the traveling standards were stable during this period.

The travelling standards were circulated among six participating laboratories without any incident that

required any return to the pilot laboratory to re measure the travelling standards.

7. REFERENCE VALUES

7.1 The reference values for this comparison were determined for the CENAM with an expanded uncertainty calculated as is described in 7.6.1, 7.6.2 and 7.3.

7.2 Before circulating the travelling standards of 2 kg, 1 kg, 200 g, 50 g and 1 g their volumes were determined at the CENAM density laboratory. The assumed density of the 200 mg weight was 7 950 kg m⁻³ as given by the manufacturer.

7.3 The expanded uncertainties of the reference values were obtained, according to GUM-1995 [4], as the combined standard uncertainties multiplied by the coverage factor *k* = 2. The expanded uncertainty corresponds to a coverage probability of approximately 95%. The uncertainty was formed from the uncertainty of measurement of the reference standard used, the weighing process and the air buoyancy correction. The uncertainty component due to long-term changes was negligible.

7.4 Table 3 gives the mass changes |Δ*m*| between the two re-calibrations by the pilot laboratory and the resulting drift uncertainties calculated by means of equation c):

Table 3. Changes in mass of the travelling standards

Nominal Value	Reference uncertainty (<i>k</i> =2)	Δ <i>m</i>	Drift uncertainty (<i>k</i> =2)
2 kg	0,08 mg	0,05 mg	0,03 mg
1 kg	0,03 mg	0,00 mg	0,00 mg
200 g	0,012 mg	0,020 mg	0,012 mg
50 g	0,006 mg	0,002 mg	0,001 mg
1 g	0,002 mg	0,000 mg	0,000 mg
200 mg	0,001 2 mg	0,000 4 mg	0,000 2 mg

7.5 The instability of the travelling standards was taken into account in the calculation of the reference values and was included in the *E_n*[5] value as an additional uncertainty component.

7.6 Therefore, the conventional mass values of the participants have to be linked to a CENAM's reference standards using the rules:

7.6.1 If two consecutive determinations of reference values are within the limits of the reference uncertainty, their mean value is used by all participants because we do not know when the value changed.

7.6.2 If two reference values *m*₁ and *m*₂, were at times *t*₁ and *t*₂, differ significantly, a linear drift is assumed and for a participant *i*, measuring at time *t*_{*i*} the mass *m*_{PL,*i*} was interpolated using the following equation.

$$m_{PL,i} = m_{PL,1} + \left(m_{PL,2} - m_{PL,1} \right) \frac{t_i - t_1}{t_2 - t_1} \quad \text{a)}$$

8. RESULTS OF PARTICIPATING LABORATORIES

8.1 The results were sent directly to the pilot laboratory (CENAM).

8.2 The results of the measurements are shown in the tabular form, see table 4 to table 9 and as a graphical representation, see 1 to 6. The *E_n* value [4] is obtained from the following expression.

$$E_n = \frac{m_A - m_{PL}}{\sqrt{U_A^2 + U_{PL}^2 + U_d^2}} \quad \text{b)}$$

*m*_{PL} and *U*_{PL} are the conventional mass value and uncertainty associated with the pilot laboratory are while *m*_A and *U*_A are the conventional mass value and the uncertainty associated with the participating laboratories.

A drift uncertainty *U_d* for the mass instability of the travelling standards is taken into account with the following equation:

$$U_d = k \sqrt{\left(\frac{m_{PL,2} - m_{PL,1}}{2\sqrt{3}} \right)^2} \quad \text{c)}$$

8.3 In the tables from 10 to 15 are included the *E_n* values for all the travelling standards and for all the participants, including the pilot laboratory. The following rules were considered.

8.4 Participant A and pilot laboratory PL.

The value E_n is calculated according to equation b):
 the measurements are considered as uncorrelated, in this case the denominator of (b) the equation is:

$$\sqrt{\left(U_A^2 + U_{PL}^2 + \frac{(m_{PL,2} - m_{PL,1})^2}{3} \right)} \quad \text{d)}$$

8.5 Participant A and B from the same petal.

The value E_n is calculated according to equation b):
 the measurements are considered as uncorrelated in this case the denominator from the equation b) is:

$$\sqrt{\left(U_A^2 + U_B^2 + U_{PL}^2 + \frac{(m_{PL,2} - m_{PL,1})^2}{3} \right)} \quad \text{e)}$$

Table 4:- Results for the 2 kg standard

Laboratory	Reference value $m_{PL}-m_n$ (mg)	Laboratory value m_l-m_n (mg)	Laboratory uncertainty U_l (mg)
CENAM		- 0,20	0,08
BSJ	- 0,225	- 0,18	2,38
LACOMET	- 0,225	- 0,24	1,46
IBMETRO	- 0,225	- 0,90	3,00
INDECOPI	- 0,225	+ 0,50	1,00
INTN	- 0,225	+ 0,20	3,00
CESMEC	- 0,225	- 0,20	1,00
CENAM		- 0,25	0,08

Table 5:- Results for the 1 kg standard

Laboratory	Reference value $m_{PL}-m_n$ (mg)	Laboratory value m_l-m_n (mg)	Laboratory uncertainty U_l (mg)
CENAM		- 0,16	0,03
BSJ	- 0,16	- 0,09	1,28
LACOMET	- 0,16	- 0,214	0,046
IBMETRO	- 0,16	- 0,128	0,154
INDECOPI	- 0,16	- 0,20	0,26
INTN	- 0,16	- 0,1	1,6
CESMEC	- 0,16	- 0,14	0,50
CENAM		- 0,16	0,03

Table 6:- Results for the 200 g standard

Laboratory	Reference value $m_{PL}-m_n$ (mg)	Laboratory value m_l-m_n (mg)	Laboratory uncertainty U_l (mg)
CENAM		- 0,365	0,012
BSJ	- 0,363	- 0,37	0,24
LACOMET	- 0,361	- 0,367 9	0,009 6
IBMETRO	- 0,356	- 0,381	0,034
INDECOPI	- 0,354	- 0,36	0,05
INTN	- 0,352	- 0,30	0,30
CESMEC	- 0,349	- 0,32	0,10
CENAM		- 0,345	0,012

Table 7:- Results for the 50 g standard

Laboratory	Reference value $m_{PL}-m_n$ (mg)	Laboratory value m_l-m_n (mg)	Laboratory uncertainty U_l (mg)
CENAM		- 0,061	0,006
BSJ	- 0,060	- 0,08	0,10
LACOMET	- 0,060	- 0,063 8	0,014 2
IBMETRO	- 0,060	- 0,052	0,016
INDECOPI	- 0,060	- 0,062	0,016
INTN	- 0,060	- 0,08	0,10
CESMEC	- 0,060	- 0,065	0,030
CENAM		- 0,059	0,006

Table 8:- Results for the 1 g standard

Laboratory	Reference value $m_{PL}-m_n$ (mg)	Laboratory value m_l-m_n (mg)	Laboratory uncertainty U_l (mg)
CENAM		+ 0,028	0,002
BSJ	+ 0,028	+ 0,030	0,040
LACOMET	+ 0,028	+ 0,025 9	0,001 5
IBMETRO	+ 0,028	+ 0,026 4	0,003 2
INDECOPI	+ 0,028	+ 0,027	0,005
INTN	+ 0,028	+ 0,027	0,030
CESMEC	+ 0,028	+ 0,027	0,010
CENAM		+ 0,028	0,002

Table 9:- Results of the standard of 200 mg

Laboratory	Reference value $m_{PL}-m_n$ (mg)	Laboratory value m_l-m_n (mg)	Laboratory uncertainty U_l (mg)
CENAM		+ 0,000 7	0,001 2
BSJ	+ 0,000 55	- 0,002	0,020
LACOMET	+ 0,000 55	+ 0,000 08	0,000 54
IBMETRO	+ 0,000 55	- 0,004 4	0,002 2
INDECOPI	+ 0,000 55	+ 0,000 9	0,003 2
INTN	+ 0,000 55	+ 0,003	0,020
CESMEC	+ 0,000 55	+ 0,001	0,006
CENAM		+ 0,000 4	0,001 2

Where:

m_n is the nominal value of the travelling standards
 m_l is the laboratory value of the travelling standards
 m_{PL} is the reference value of the travelling standards

9. CONCLUSIONS

Of the 36 measurements results for the mass of the travelling standards, were used for calculating the respective E_n value between participants, see tables 10 to 15 of which one are greater than one respect to pilot laboratory and two are greater than one between them, these values were calculated using the d) and e) formulas. The degree of agreement among the participants seen to be excellent, in other words, the mass measurements carried out among SIM region members do not differ significantly.

The uncertainties of each participant are plotted in graphical representation, see graphics 1 to 6 using the d) formula.

The names of the participating laboratories were included in the report as an agreement between them.

ACKNOWLEDGMENT

Some of the participant institutions wish to acknowledge to their colleagues who they carried out the mass measurements:

Leticia Luján Solís
Raúl Hernández

CENAM
CESMEC

Sheldon Walker
Oscar Andrey Herrera Sancho


BSJ
LACOMET

REFERENCES


- [1] OIML R111, Weights of classes E₁, E₂, F₁, F₂, M₁, M₂, M₃, 1994.
- [2] Norma Oficial Mexicana NOM-038-SCFI-2000, Pesa de clase de exactitud E₁, E₂, F₁, F₂, M₁, M₂, M₃.
- [3] Guidelines for CIPM key comparisons <http://www.bipm.fr/utis/en/pdf/guidelines.pdf>
- [4] Guide to the Expression of Uncertainty in Measurement, International Organization for Standardization, Geneva, Switzerland, 1995.
- [5] W. Wöger, "Remarks on the E_n -criterion Used in Measurement Comparison, Internationale Zusammenarbeit.PTB-Mitteilungen 1999. Pages 24-27.

APPENDIX A – GRAPHICS


Graphic 1: The difference between the laboratory value and the reference value: 2 kg




Graphic 2: The difference between the laboratory value and the reference value: 1 kg




Graphic 3: The difference between the laboratory value and the reference value: 200 g




Graphic 4: The difference between the laboratory value and the reference value: 50 g



Graphic 6: The difference between the laboratory value and the reference value: 200 mg



Graphic 5: The difference between the laboratory value and the reference value: 1g



APPENDIX B – E_n VALUES

Table 10: The E_n value for 2 kg

2 kg	CENAM	BSJ	LACOMET	IBMETRO	INDECOPI	INTN	CESMEC
CENAM		0.02	-0.01	-0.23	0.72	0.14	0.02
BSJ	-0.02		-0.02	-0.19	0.26	0.10	-0.01
LACOMET	0.01	0.02		-0.20	0.42	0.13	0.02
IBMETRO	0.23	0.19	0.20		0.44	0.26	0.22
INDECOPI	-0.72	-0.26	-0.42	-0.44		-0.09	-0.49
INTN	-0.14	-0.10	-0.13	-0.26	0.09		-0.13
CESMEC	-0.02	0.01	-0.02	-0.22	0.49	0.13	

Table 11: The E_n value for 1 kg

1 kg	CENAM	BSJ	LACOMET	IBMETRO	INDECOPI	INTN	CESMEC
CENAM		0.05	-0.98	0.20	-0.15	0.04	0.04
BSJ	-0.05		-0.10	-0.03	-0.08	0.00	-0.04
LACOMET	0.98	0.10		0.53	0.05	0.07	0.15
IBMETRO	-0.20	0.03	-0.53		-0.24	0.02	-0.02
INDECOPI	0.15	0.08	-0.05	0.24		0.06	0.11
INTN	-0.04	0.00	-0.07	-0.02	-0.06		-0.02
CESMEC	-0.04	0.04	-0.15	0.02	-0.11	0.02	

Table 12: The E_n value for 200 g

200 g	CENAM	BSJ	LACOMET	IBMETRO	INDECOPI	INTN	CESMEC
CENAM		-0.03	-0.38	-0.66	-0.12	0.17	0.29
BSJ	0.03		0.01	-0.05	0.04	0.18	0.19
LACOMET	0.38	-0.01		-0.33	0.15	0.23	0.47
IBMETRO	0.66	0.05	0.33		0.33	0.27	0.57
INDECOPI	0.12	-0.04	-0.15	-0.34		0.20	0.35
INTN	-0.17	-0.18	-0.23	-0.27	-0.20		-0.06
CESMEC	-0.29	-0.19	-0.47	-0.57	-0.35	0.06	

Table 13: The E_n value for 50 g

50 g	CENAM	BSJ	LACOMET	IBMETRO	INDECOPI	INTN	CESMEC
CENAM		-0.20	-0.25	0.47	-0.12	-0.20	-0.16
BSJ	0.20		0.16	0.28	0.18	0.00	0.14
LACOMET	0.25	-0.16		0.53	0.08	-0.16	-0.04
IBMETRO	-0.47	-0.28	-0.53		-0.43	-0.28	-0.38
INDECOPI	0.12	-0.18	-0.08	0.43		-0.18	-0.09
INTN	0.20	0.00	0.16	0.28	0.18		0.14
CESMEC	0.16	-0.14	0.04	0.38	0.09	-0.14	

Table 14: The E_n value for 1 g

1 g	CENAM	BSJ	LACOMET	IBMETRO	INDECOPI	INTN	CESMEC
CENAM		0.04	-0.94	-0.48	-0.23	-0.01	-0.12
BSJ	-0.04		-0.10	-0.09	-0.07	-0.03	-0.07
LACOMET	0.94	0.10		0.13	0.21	0.01	0.11
IBMETRO	0.48	0.09	-0.13		0.10	0.01	0.06
INDECOPI	0.23	0.07	-0.21	-0.10		0.00	0.00
INTN	0.01	0.03	-0.01	-0.01	0.00		0.00
CESMEC	0.12	0.07	-0.11	-0.06	0.00	0.00	

Table 15: The E_n value for 200 mg

200 mg	CENAM	BSJ	LACOMET	IBMETRO	INDECOPI	INTN	CESMEC
CENAM		-0.13	-0.34	-1.96	0.11	0.12	0.08
BSJ	0.13		0.10	-0.12	0.14	0.18	0.14
LACOMET	0.34	-0.10		-1.75	0.24	0.15	0.15
IBMETRO	1.96	0.12	1.75		1.30	0.37	0.83
INDECOPI	-0.11	-0.14	-0.24	-1.30		0.10	0.01
INTN	-0.12	-0.18	-0.15	-0.37	-0.10		-0.10
CESMEC	-0.08	-0.14	-0.15	-0.83	-0.01	0.10	